



## Non PVC Fluid Path

6" Macrobore extension set with pinch clamp and male luer-lock with non-bonded K-Zero needleless connector

## FOR PARENTERAL FLUIDS

### CAUTIONS:

- Do not use if package is damaged and/or protection caps are detached. Keep dry. Store at controlled room temperature. Keep away from sunlight.

  • Use aseptic technique.
- · Single use only.

### MORE INFORMATION:

- · Length is approximate
- Approximate contained volume 1.2 mL
- Gravity flow rate 4,980 mL/hr
- Sterile and nonpyrogenic fluid path
   Fluid Path Material: Non-PVC TPU (Not made with DEHP) / PE (end cap) / PC (luer locks)

## CONTRAINDICATIONS:

· Materials of the extension set does not provide light protection.

### PRIMING:

### Priming must be completed before connecting to patient.

- · Connect a fully primed infusion set or syringe to extension set female luer
- Connect K-Zero to primary set
- · Tighten all connections
- · Prime extension set before attaching to patient
- Clamp set
- Swab patient's access site or other luer connection before connecting to patient per facility protocol
- Remove end cap and attach to patient luer

## ATTACHING TO NEEDLELESS CONNECTORS:

- · Swab surface of the needleless connector prior to and after every use according to CDC guidelines or facility protocol. Use luer lock connections only. If using luer slip, do not leave unattended.
- Do not use sharp needle or blunt cannula with the needleless connectors.

# SET REMOVAL:

• Stop infusion. Close pinch clamp, disconnect the set from IV access and container in accordance with facility protocol. Replace per facility protocol or CDC guidelines. Discard biohazard waste according to facility protocol.

Manufactured For: Fresenius Kabi AG Else-Kröner-Str.1

61352 Bad Homburg, Germany www.fresenius-kabi.com

Distributed in USA by: Fresenius Kabi USA, LLC Three Corporate Drive Lake Zurich, Illinois 60047 www.fresenius-kabi.us Tel.: 1-800-933-6925

N. Code No. : HR/Drugs 737-B(H)

10.5933.02 10.5520.00